



**REF:** TBA

**Product Name:** pH Buffer

**Specification:** 2mL, 40pcs/box

**Expiry Date:** 2 years.

**Intended use:**

Adjust the pH value during the Endotoxin detection testing.

**Procedures:**

Using the pH buffer to adjust the pH value of the sample to 5-6, also refer to the IFU of TAL or Endotoxin Assay Kit.

**Precautions:**

Don't use pH buffer to constitute TAL, it is corrosive.



Zhanjiang Bokang Marine Biological co., Ltd.

#188, RuiYunNan Rd., MaZhang district, Zhan Jiang, Guang Dong, China

Zipcode: 524094